

Distal Gradient Tracking (DGT) for Ion Beam Radiation Therapy

View U.S. Patent No. 7,977,657 in PDF format.

WARF: P07280US

Inventors: Thomas (Rock) Mackie, Ryan Flynn

The Wisconsin Alumni Research Foundation (WARF) is seeking commercial partners interested in developing an ion radiation treatment system that places the Bragg peak according to a prescribed dose plan to better control radiation dose to the tumor area.

Overview

External beam radiation therapy is a tumor treatment technique that directs one or more high-energy radiation beams to the tumor. Advanced external beam radiation systems treat tumors with multiple X-ray fan beams that can be rotated around the patient. Each beam consists of individual "beamlets" whose intensity can be controlled to allow for the treatment of complex tumor shapes. However, X-ray energy still is deposited to all tissue along the entire path of the beam up to the exit point.

An alternate beam radiation therapy technique involves the use of ions such as protons to treat tumors. This technique dramatically reduces the radiation dose to healthy tissue when compared with X-ray radiation therapy using photons. The dose intensity from protons is not uniform along the beam path and rises to a "Bragg peak" near a point where the proton beam stops completely. Controlling the placement of the Bragg peak so it is located on the tumor reduces the amount of radiation delivered to the patient's healthy tissue. Unlike X-ray radiation therapy, ions allow for separate control of the total dose of radiation (intensity) and distance the Bragg peak occurs (range).

Digital Edge Tracking (DET) is a proton beamlet placement process in which the Bragg peak occurs at the point that the beam axis exits the tumor. This process applies a homogeneous dose of ion radiation to the target tumor. However, the treatment of certain tumors can benefit from a non-homogeneous dose tailored to the specific tumor.

The Invention

UW-Madison researchers have developed an ion radiation treatment system that places the Bragg peak according to a prescribed dose plan to impart non-homogeneous radiation doses to the tumor area. Basically, the Bragg peak is placed at a point where the prescribed dose value falls under a certain level in a dose gradient. Each gradient is specific to the angle that the beam is being applied to, and thus is tailored to the specific tumor.

The system consists of an ion beam source and a range controller. A separate beam controller executes a stored radiation plan in conjunction with a dose plan that defines the region where the dose is to be applied to determine the beam gradients for every angle at which the ion beam is applied. The beam controller sends control signals to the range controller to position the Bragg peaks according to the gradients.

Applications

e cookies on this site to enhance your experience and improve our marketing efforts. By continuing to browse without changing your browser settings to block or delete

External ion beam radiation therapyofor treatment of intermore lated technologies on your device. See our privacy policy



Key Benefits

- · Decreases the amount of proton beamlets needed
- · Allows for tailored dose prescriptions inside the tumor
- · Increases treatment speed

Tech Fields

• Radiation Therapy : External beam therapy

For current licensing status, please contact Jeanine Burmania at jeanine@warf.org or 608-960-9846

We use cookies on this site to enhance your experience and improve our marketing efforts. By continuing to browse without changing your browser settings to block or delete cookies, you agree to the storing of cookies and related technologies on your device. See our privacy policy

